# Manufacturer Disclosure Statement for Medical Device Security -- MDS2

Clarius Mobile Health Clarius Ultrasound App 51-10-00002 22-Oct-2020

Question ID	Question		See note
DOC-1	Manufacturer Name	Clarius Mobile Health	_
DOC-2	Device Description	Clarius Ultrasound Scanner	_
DOC-3	Device Model	Clarius Ultrasound App	_
DOC-4	Document ID  Manufacturer Contact Information	51-10-00002 support@clarius.com	_
DOC-5	Manufacturer Contact Information	support@ciarius.com	_
		The Clarius Ultrasound App runs on a user provided mobile device (iOS or Android based). The App acquires ultrasound images from the Clarius scanner. Patient information can be added to the App. Once an examination is complete, patient	
		information and images can be uploaded to the Clarius Cloud or to a	
	Intended use of device in network-connected	customer's PACS where they may be	
DOC-6	environment:	stored permanently.	_
DOC-7	Document Release Date	2020-10-22	_
DOC 0	Coordinated Vulnerability Disclosure: Does the manufacturer have a vulnerability disclosure program for		
DOC-8	this device? ISAO: Is the manufacturer part of an Information Sharing	No	_
DOC-9	and Analysis Organization?	No	
	Diagram: Is a network or data flow diagram available that		
	indicates connections to other system components or		
DOC-10	expected external resources?		See Note 1
DOC-11	SaMD: Is the device Software as a Medical Device (i.e. software-only, no hardware)?	No	
DOC-11.1	Does the SaMD contain an operating system?	N/A	_
	Does the SaMD rely on an owner/operator provided	•	_
DOC-11.2	operating system?	N/A	_
	Is the SaMD hosted by the manufacturer?		
DOC-11.3 DOC-11.4	Is the SaMD hosted by the customer?	N/A N/A	
	MANAGEMENT OF PERSONALLY IDENTIFIABLE INFORMATION	Yes, No, N/A, or See Note	Note#
	Can this device display, transmit, store, or modify		
MPII-1	personally identifiable information (e.g. electronic Protected Health Information (ePHI))?	Yes	
IAIL II-T	Does the device maintain personally identifiable	16	_
MPII-2	information?	Yes	
MPII-2.1	Does the device maintain personally identifiable information temporarily in volatile memory (i.e., until cleared by power-off or reset)?	Yes	
	Does the device store personally identifiable information		
MPII-2.2	persistently on internal media? Is personally identifiable information preserved in the	Yes	See Note 2
MPII-2.3	device's non-volatile memory until explicitly erased?  Does the device store personally identifiable information	Yes	See Note 2
MPII-2.4	in a database?	No	_
MPII-2.5	Does the device allow configuration to automatically delete local personally identifiable information after it is stored to a long term solution?	Yes	See Note 2
MDU 2.6	Does the device import/export personally identifiable information with other systems (e.g., a wearable monitoring device might export personally identifiable information to a correct?)	Ven	Con Note 2
MPII-2.6	information to a server)?  Does the device maintain personally identifiable information when powered off, or during power service	Yes	See Note 3
MPII-2.7	interruptions?  Does the device allow the internal media to be removed	Yes	-
MPII-2.8	by a service technician (e.g., for separate destruction or customer retention)?	N/A	

	Does the device allow personally identifiable		
	information records be stored in a separate location		
	from the device's operating system (i.e. secondary		
	internal drive, alternate drive partition, or remote		
MPII-2.9	storage location)?	N/A	
	Does the device have mechanisms used for the		
	transmitting, importing/exporting of personally		
MPII-3	identifiable information?	Yes	See Note 3
	Does the device display personally identifiable		
MPII-3.1	information (e.g., video display, etc.)?	Yes	See Note 4
	Does the device generate hardcopy reports or images		
MPII-3.2	containing personally identifiable information?	Yes	See Note 5
	Does the device retrieve personally identifiable		
	information from or record personally identifiable		
	information to removable media (e.g., removable-HDD,		
	USB memory, DVD-R/RW,CD-R/RW, tape, CF/SD card,		
MPII-3.3	memory stick, etc.)?	No	
	Does the device transmit/receive or import/export		_
	personally identifiable information via dedicated cable		
MPII-3.4	connection (e.g., RS-232, RS-423, USB, FireWire, etc.)?	No	
	Does the device transmit/receive personally identifiable		_
	information via a wired network connection (e.g., RJ45,		
MPII-3.5	fiber optic, etc.)?	N/A	
		,	<del></del>
	Does the device transmit/receive personally identifiable		
	information via a wireless network connection (e.g.,		
MPII-3.6	WiFi, Bluetooth, NFC, infrared, cellular, etc.)?	Yes	See Note 3
	Does the device transmit/receive personally identifiable		
MPII-3.7	information over an external network (e.g., Internet)?	Yes	See Note 3
	Does the device import personally identifiable		
MPII-3.8	information via scanning a document?	Yes	See Note 6
	Does the device transmit/receive personally identifiable		
MPII-3.9	information via a proprietary protocol?	Yes	See Note 7
	Does the device use any other mechanism to transmit,		
MPII-3.10	import or export personally identifiable information?	No	
Management of Priva			_

## AUTOMATIC LOGOFF (ALOF)

The device's ability to prevent access and misuse by unauthorized users if device is left idle for a period of time.

Can the device be configured to force reauthorization of logged-in user(s) after a predetermined length of		
inactivity (e.g., auto-logoff, session lock, password		
protected screen saver)?	See Notes	See Note 8
Is the length of inactivity time before auto-logoff/screen		
lock user or administrator configurable?	See Notes	See Note 9

#### AUDIT CONTROLS (AUDT)

ALOF-1

ALOF-2

	The ability to reliably audit activity on the device.		
	Can the medical device create additional audit logs or		
AUDT-1	reports beyond standard operating system logs?	No	
AUDT-1.1	Does the audit log record a USER ID?	N/A	
	Does other personally identifiable information exist in		
AUDT-1.2	the audit trail?	N/A	
	Are events recorded in an audit log? If yes, indicate which		
AUDT-2	of the following events are recorded in the audit log:	No	
AUDT-2.1	Successful login/logout attempts?	N/A	
AUDT-2.2	Unsuccessful login/logout attempts?	N/A	
AUDT-2.3	Modification of user privileges?	N/A	
AUDT-2.4	Creation/modification/deletion of users?	N/A	
AUDT-2.5	Presentation of clinical or PII data (e.g. display, print)?	N/A	
AUDT-2.6	Creation/modification/deletion of data?	N/A	
	Import/export of data from removable media (e.g. USB		
AUDT-2.7	drive, external hard drive, DVD)?	N/A	
	Receipt/transmission of data or commands over a		
AUDT-2.8	network or point-to-point connection?	N/A	

AUDT-2.8.1	Remote or on-site support?	N/A
AUD1-2.6.1	··	
AUDT 2 0 2	Application Programming Interface (API) and similar	N/A
AUDT-2.8.2	activity?	N/A
AUDT-2.9	Emergency access?	N/A
AUDT-2.10	Other events (e.g., software updates)?	N/A
AUDT-2.11	Is the audit capability documented in more detail?	N/A
	Can the owner/operator define or select which events are	
AUDT-3	recorded in the audit log?	No
	Is a list of data attributes that are captured in the audit	
AUDT-4	log for an event available?	No
AUDT-4.1	Does the audit log record date/time?	N/A
	Can date and time be synchronized by Network Time	<del>-</del>
AUDT-4.1.1	Protocol (NTP) or equivalent time source?	N/A
AUDT-5	Can audit log content be exported?	No —
AUDT-5.1	Via physical media?	N/A
	Via IHE Audit Trail and Node Authentication (ATNA)	<del>-</del>
AUDT-5.2	profile to SIEM?	N/A
	Via Other communications (e.g., external service device,	<del>-</del>
AUDT-5.3	mobile applications)?	N/A
AUDT-5.4	Are audit logs encrypted in transit or on storage media?	N/A
	Can audit logs be monitored/reviewed by	
AUDT-6	owner/operator?	No
AUDT-7	Are audit logs protected from modification?	No
AUDT-7.1	Are audit logs protected from access?	N/A
AUDT-8	Can audit logs be analyzed by the device?	No

#### **AUTHORIZATION (AUTH)**

The ability of the device to determine the authorization of users.

	Does the device prevent access to unauthorized users		
AUTH-1	through user login requirements or other mechanism?	Yes	_
	Can the device be configured to use federated credentials		
	management of users for authorization (e.g., LDAP,		
AUTH-1.1	OAuth)?	No	_
	Can the customer push group policies to the device (e.g.,		
AUTH-1.2	Active Directory)?	No	_
	Are any special groups, organizational units, or group		
AUTH-1.3	policies required?	No	_
	Can users be assigned different privilege levels based on		
AUTH-2	'role' (e.g., user, administrator, and/or service, etc.)?	Yes	See Note 10
	Can the device owner/operator grant themselves		
	unrestricted administrative privileges (e.g., access		
	operating system or application via local root or		
AUTH-3	administrator account)?	No	_
	Does the device authorize or control all API access		
AUTH-4	requests?	N/A	_
	Does the device run in a restricted access mode, or 'kiosk		
AUTH-5	mode', by default?	N/A	_

## CYBER SECURITY PRODUCT UPGRADES (CSUP)

The ability of on-site service staff, remote service staff, or authorized customer staff to install/upgrade device's security patches.

CSUP-1	Does the device contain any software or firmware which may require security updates during its operational life, either from the device manufacturer or from a third-party manufacturer of the software/firmware? If no, answer "N/A" to questions in this section.	Yes	See Note 11
C3UP-1		res	See Note 11
CSUP-2	Does the device contain an Operating System? If yes, complete 2.1-2.4.	Yes	See Note 11
CSUP-2.1	Does the device documentation provide instructions for owner/operator installation of patches or software updates?	Yes	
CSUP-2.1	·	Yes	_
CSUP-2.2	Does the device require vendor or vendor-authorized service to install patches or software updates?	N/A	_
CSUP-2.3	Does the device have the capability to receive remote installation of patches or software updates?	No	_

	Does the medical device manufacturer allow security	
	updates from any third-party manufacturers (e.g.,	
	Microsoft) to be installed without approval from the	
CSUP-2.4	manufacturer?	N/A
	Does the device contain Drivers and Firmware? If yes,	
CSUP-3	complete 3.1-3.4.	Yes
	Does the device documentation provide instructions for	
	owner/operator installation of patches or software	
CSUP-3.1	updates?	Yes See Note 11
	Does the device require vendor or vendor-authorized	
CSUP-3.2	service to install patches or software updates?	No
	Does the device have the capability to receive remote	
CSUP-3.3	installation of patches or software updates?	N/A
	Does the medical device manufacturer allow security	
	updates from any third-party manufacturers (e.g.,	
	Microsoft) to be installed without approval from the	
CSUP-3.4	manufacturer?	N/A
	Does the device contain Anti-Malware Software? If yes,	
CSUP-4	complete 4.1-4.4.	No
	Does the device documentation provide instructions for	
	owner/operator installation of patches or software	
CSUP-4.1	updates?	N/A
	Does the device require vendor or vendor-authorized	
CSUP-4.2	service to install patches or software updates?	N/A
	Does the device have the capability to receive remote	
CSUP-4.3	installation of patches or software updates?	N/A
	Does the medical device manufacturer allow security	
	updates from any third-party manufacturers (e.g.,	
	Microsoft) to be installed without approval from the	
CSUP-4.4	manufacturer?	N/A
	Does the device contain Non-Operating System	_
	commercial off-the-shelf components? If yes, complete	
CSUP-5	5.1-5.4.	No
	Does the device documentation provide instructions for	_
	owner/operator installation of patches or software	
CSUP-5.1	updates?	N/A
	Does the device require vendor or vendor-authorized	_
CSUP-5.2	service to install patches or software updates?	N/A
	Does the device have the capability to receive remote	_
CSUP-5.3	installation of patches or software updates?	N/A
	Does the medical device manufacturer allow security	
	updates from any third-party manufacturers (e.g.,	
	Microsoft) to be installed without approval from the	
CSUP-5.4	manufacturer?	N/A
	Does the device contain other software components	
	(e.g., asset management software, license management)?	
	If yes, please provide details or refernce in notes and	
CSUP-6	complete 6.1-6.4.	No
	Does the device documentation provide instructions for	
	= = = = action action accuration provide matriculation	
	owner/operator installation of patches or software	
CSUP-6.1	owner/operator installation of patches or software updates?	N/A
CSUP-6.1	updates?	N/A
	updates?  Does the device require vendor or vendor-authorized	_
CSUP-6.1	updates?  Does the device require vendor or vendor-authorized service to install patches or software updates?	N/A N/A
CSUP-6.2	updates?  Does the device require vendor or vendor-authorized service to install patches or software updates?  Does the device have the capability to receive remote	N/A
	updates?  Does the device require vendor or vendor-authorized service to install patches or software updates?  Does the device have the capability to receive remote installation of patches or software updates?	_
CSUP-6.2	updates?  Does the device require vendor or vendor-authorized service to install patches or software updates?  Does the device have the capability to receive remote installation of patches or software updates?  Does the medical device manufacturer allow security	N/A
CSUP-6.2	updates?  Does the device require vendor or vendor-authorized service to install patches or software updates?  Does the device have the capability to receive remote installation of patches or software updates?  Does the medical device manufacturer allow security updates from any third-party manufacturers (e.g.,	N/A
CSUP-6.2 CSUP-6.3	updates?  Does the device require vendor or vendor-authorized service to install patches or software updates?  Does the device have the capability to receive remote installation of patches or software updates?  Does the medical device manufacturer allow security updates from any third-party manufacturers (e.g., Microsoft) to be installed without approval from the	N/A
CSUP-6.2 CSUP-6.3	updates?  Does the device require vendor or vendor-authorized service to install patches or software updates?  Does the device have the capability to receive remote installation of patches or software updates?  Does the medical device manufacturer allow security updates from any third-party manufacturers (e.g., Microsoft) to be installed without approval from the manufacturer?	N/A
CSUP-6.2 CSUP-6.3 CSUP-6.4	updates?  Does the device require vendor or vendor-authorized service to install patches or software updates?  Does the device have the capability to receive remote installation of patches or software updates?  Does the medical device manufacturer allow security updates from any third-party manufacturers (e.g., Microsoft) to be installed without approval from the manufacturer?  Does the manufacturer notify the customer when	N/A  N/A
CSUP-6.2 CSUP-6.3 CSUP-6.4	updates?  Does the device require vendor or vendor-authorized service to install patches or software updates?  Does the device have the capability to receive remote installation of patches or software updates?  Does the medical device manufacturer allow security updates from any third-party manufacturers (e.g., Microsoft) to be installed without approval from the manufacturer?  Does the manufacturer notify the customer when updates are approved for installation?	N/A
CSUP-6.3 CSUP-6.4 CSUP-7	updates?  Does the device require vendor or vendor-authorized service to install patches or software updates?  Does the device have the capability to receive remote installation of patches or software updates?  Does the medical device manufacturer allow security updates from any third-party manufacturers (e.g., Microsoft) to be installed without approval from the manufacturer?  Does the manufacturer notify the customer when updates are approved for installation?  Does the device perform automatic installation of	N/A  N/A  N/A  N/A
CSUP-6.2	updates?  Does the device require vendor or vendor-authorized service to install patches or software updates?  Does the device have the capability to receive remote installation of patches or software updates?  Does the medical device manufacturer allow security updates from any third-party manufacturers (e.g., Microsoft) to be installed without approval from the manufacturer?  Does the manufacturer notify the customer when updates are approved for installation?  Does the device perform automatic installation of software updates?	N/A  N/A
CSUP-6.2 CSUP-6.3 CSUP-6.4 CSUP-7 CSUP-8	updates?  Does the device require vendor or vendor-authorized service to install patches or software updates?  Does the device have the capability to receive remote installation of patches or software updates?  Does the medical device manufacturer allow security updates from any third-party manufacturers (e.g., Microsoft) to be installed without approval from the manufacturer?  Does the manufacturer notify the customer when updates are approved for installation?  Does the device perform automatic installation of software updates?  Does the manufacturer have an approved list of third-	N/A  N/A  N/A  N/A  N/A  See Notes
CSUP-6.2 CSUP-6.3 CSUP-6.4 CSUP-7	updates?  Does the device require vendor or vendor-authorized service to install patches or software updates?  Does the device have the capability to receive remote installation of patches or software updates?  Does the medical device manufacturer allow security updates from any third-party manufacturers (e.g., Microsoft) to be installed without approval from the manufacturer?  Does the manufacturer notify the customer when updates are approved for installation?  Does the device perform automatic installation of software updates?	N/A  N/A  N/A  N/A
CSUP-6.2 CSUP-6.3 CSUP-6.4 CSUP-7 CSUP-8	updates?  Does the device require vendor or vendor-authorized service to install patches or software updates?  Does the device have the capability to receive remote installation of patches or software updates?  Does the medical device manufacturer allow security updates from any third-party manufacturers (e.g., Microsoft) to be installed without approval from the manufacturer?  Does the manufacturer notify the customer when updates are approved for installation?  Does the device perform automatic installation of software updates?  Does the manufacturer have an approved list of third-party software that can be installed on the device?	N/A  N/A  N/A  N/A  N/A  See Notes
CSUP-6.2 CSUP-6.3 CSUP-6.4 CSUP-7 CSUP-8	updates?  Does the device require vendor or vendor-authorized service to install patches or software updates?  Does the device have the capability to receive remote installation of patches or software updates?  Does the medical device manufacturer allow security updates from any third-party manufacturers (e.g., Microsoft) to be installed without approval from the manufacturer?  Does the manufacturer notify the customer when updates are approved for installation?  Does the device perform automatic installation of software updates?  Does the manufacturer have an approved list of third-party software that can be installed on the device?  Can the owner/operator install manufacturer-approved	N/A  N/A  N/A  N/A  N/A  See Notes
CSUP-6.2 CSUP-6.3 CSUP-6.4 CSUP-7 CSUP-8	updates?  Does the device require vendor or vendor-authorized service to install patches or software updates?  Does the device have the capability to receive remote installation of patches or software updates?  Does the medical device manufacturer allow security updates from any third-party manufacturers (e.g., Microsoft) to be installed without approval from the manufacturer?  Does the manufacturer notify the customer when updates are approved for installation?  Does the device perform automatic installation of software updates?  Does the manufacturer have an approved list of third-party software that can be installed on the device?	N/A  N/A  N/A  N/A  N/A  See Notes

CSUP-11	Does the manufacturer have a process in place to assess device vulnerabilities and updates?	N/A
	Does the manufacturer provide customers with review	
CSUP-11.1	and approval status of updates?	N/A
CSUP-11.2	Is there an update review cycle for the device?	N/A

#### **HEALTH DATA DE-IDENTIFICATION (DIDT)**

The ability of the device to directly remove information

that allows identification of a person.

Does the device provide an integral capability to deidentify personally identifiable information?

Does the device support de-identification profiles that
DIDT-1.1 comply with the DICOM standard for de-identification?

Yes \_\_\_

#### **DATA BACKUP AND DISASTER RECOVERY (DTBK)**

The ability to recover after damage or destruction of device data, hardware, software, or site configuration information.

Does the device maintain long term primary storage of

personally identifiable information / patient information (e.g. PACS)?

Does the device have a "factory reset" function to restore the original device settings as provided by the manufacturer?

Does the device have an integral data backup capability to removable media?

Does the device have an integral data backup capability to remote storage?

Does the device have a backup capability for system configuration information, patch restoration, and software restoration?

Does the device provide the capability to check the

Does the device provide the capability to check the DTBK-6 integrity and authenticity of a backup?

DTBK-1

DTBK-2

DTBK-3

DTBK-4

DTBK-5

EMRG-1

IGAU-1

IGAU-2

MLDP-1

MLDP-2

MLDP-2.1

No \_\_\_Yes

No \_\_\_

No N/A

No

#### **EMERGENCY ACCESS (EMRG)**

The ability of the device user to access personally identifiable information in case of a medical emergency situation that requires immediate access to stored personally identifiable information.

Does the device incorporate an emergency access (i.e. "break-glass") feature?

No \_\_\_

# HEALTH DATA INTEGRITY AND AUTHENTICITY (IGAU)

How the device ensures that the stored data on the device has not been altered or destroyed in a non-authorized manner and is from the originator.

Does the device provide data integrity checking mechanisms of stored health data (e.g., hash or digital signature)?

Does the device provide error/failure protection and

Does the device provide error/failure protection and recovery mechanisms for stored health data (e.g., RAID-5)?

# No \_\_\_\_\_

#### MALWARE DETECTION/PROTECTION (MLDP)

The ability of the device to effectively prevent, detect and remove malicious software (malware). Is the device capable of hosting executable software? Does the device support the use of anti-malware software (or other anti-malware mechanism)? Provide details or reference in notes.

Does the device include anti-malware software by default?

 N/A
 \_\_

 See Notes
 See Note 13

 N/A
 \_\_

See Note 3

N/A

N/A

No

Yes

CONN-1.2.3

CONN-1.2.4

CONN-2

CONN-3

memory devices?

the customer environment?

device?

Does the device support other physical connectivity?

Does the manufacturer provide a list of network ports and protocols that are used or may be used on the

Can the device communicate with other systems within

	Can the device communicate with other systems external		
CONN-4	to the customer environment (e.g., a service host)?	Yes	See Note 3
CONN-5	Does the device make or receive API calls?	Yes	See Note 3
	Does the device require an internet connection for its		
CONN-6	intended use?	No	See Note 14
CONN-7	Does the device support Transport Layer Security (TLS)?	Yes	_
CONN-7.1	Is TLS configurable?	No	
	Does the device provide operator control functionality		
CONN-8	from a separate device (e.g., telemedicine)?	No	_

#### PERSON AUTHENTICATION (PAUT)

The ability to configure the device to authenticate users.

	users.		
	Does the device support and enforce unique IDs and		
	passwords for all users and roles (including service		
PAUT-1	accounts)?	Yes	See Note 15
	Does the device enforce authentication of unique IDs and		
	passwords for all users and roles (including service		
PAUT-1.1	accounts)?	Yes	_
	Is the device configurable to authenticate users through		
	an external authentication service (e.g., MS Active		
PAUT-2	Directory, NDS, LDAP, OAuth, etc.)?	No	_
	Is the device configurable to lock out a user after a		
PAUT-3	certain number of unsuccessful logon attempts?	No	_
	Are all default accounts (e.g., technician service		
	accounts, administrator accounts) listed in the		
PAUT-4	documentation?	N/A	_
PAUT-5	Can all passwords be changed?	Yes	_
	Is the device configurable to enforce creation of user		
	account passwords that meet established (organization		
PAUT-6	specific) complexity rules?	Yes	_
	Does the device support account passwords that expire		
PAUT-7	periodically?	Yes	_
PAUT-8	Does the device support multi-factor authentication?	No	_
PAUT-9	Does the device support single sign-on (SSO)?	No	_
PAUT-10	Can user accounts be disabled/locked on the device?	Yes	_
PAUT-11	Does the device support biometric controls?	No	_
	Does the device support physical tokens (e.g. badge		
PAUT-12	access)?	No	_
	Does the device support group authentication (e.g.		
PAUT-13	hospital teams)?	No	_
	Does the application or device store or manage		
PAUT-14	authentication credentials?	No	_
PAUT-14.1	Are credentials stored using a secure method?	N/A	_

#### PHYSICAL LOCKS (PLOK)

Physical locks can prevent unauthorized users with physical access to the device from compromising the integrity and confidentiality of personally identifiable information stored on the device or on removable media

Is the device software only? If yes, answer "N/A" to remaining questions in this section.

Are all device components maintaining personally identifiable information (other than removable media)

PLOK-2 physically secure (i.e., cannot remove without tools)?

Are all device components maintaining personally identifiable information (other than removable media) physically secured behind an individually keyed locking device?

Does the device have an option for the customer to attach a physical lock to restrict access to removable PLOK-4 media?

No	_
See Notes	See Note 16
See Notes	See Note 16
See Notes	See Note 16

# ROADMAP FOR THIRD PARTY COMPONENTS IN DEVICE LIFE CYCLE (RDMP)

	Manufacturer's plans for security support of third-party components within the device's life cycle.		
RDMP-1	Was a secure software development process, such as ISO/IEC 27034 or IEC 62304, followed during product development?	Yes	
RDMP-2	Does the manufacturer evaluate third-party applications and software components included in the device for secure development practices?	Yes	_
RDMP-3	Does the manufacturer maintain a web page or other source of information on software support dates and updates?	No	_
RDIVIP-3	Does the manufacturer have a plan for managing third-	NO	_
RDMP-4	party component end-of-life?	Yes	_

#### SOFTWARE BILL OF MATERIALS (SBoM)

A Software Bill of Material (SBoM) lists all the software components that are incorporated into the device being described for the purpose of operational security planning by the healthcare delivery organization. This section supports controls in the RDMP section.

	section supports controls in the RDMP section.		
SBOM-1	Is the SBoM for this product available?	No	_
	Does the SBoM follow a standard or common method in		
SBOM-2	describing software components?	N/A	_
SBOM-2.1	Are the software components identified?	N/A	_
	Are the developers/manufacturers of the software		
SBOM-2.2	components identified?	N/A	_
	Are the major version numbers of the software		
SBOM-2.3	components identified?	N/A	_
SBOM-2.4	Are any additional descriptive elements identified?	N/A	_
	Does the device include a command or process method		
	available to generate a list of software components		
SBOM-3	installed on the device?	No	_
SBOM-4	Is there an update process for the SBoM?	N/A	_

#### SYSTEM AND APPLICATION HARDENING (SAHD)

 $\label{thm:continuous} The \ device's inherent \ resistance \ to \ cyber \ attacks \ and \ malware.$ 

	Is the device hardened in accordance with any industry		
SAHD-1	standards?	See Notes	See Note 17
SAHD-2	Has the device received any cybersecurity certifications?	No	_
	Does the device employ any mechanisms for software		
SAHD-3	integrity checking	Yes	_
	Does the device employ any mechanism (e.g., release-		
	specific hash key, checksums, digital signature, etc.) to		
	ensure the installed software is manufacturer-		
SAHD-3.1	authorized?	Yes	_
	Does the device employ any mechanism (e.g., release-		
	specific hash key, checksums, digital signature, etc.) to		
	ensure the software updates are the manufacturer-		
SAHD-3.2	authorized updates?	Yes	_
	Can the owner/operator perform software integrity		
	checks (i.e., verify that the system has not been modified		
SAHD-4	or tampered with)?	No	_
	Is the system configurable to allow the implementation		
	of file-level, patient level, or other types of access		
SAHD-5	controls?	N/A	_
SAHD-5.1	Does the device provide role-based access controls?	N/A	_
	Are any system or user accounts restricted or disabled by		
SAHD-6	the manufacturer at system delivery?	N/A	_
	Are any system or user accounts configurable by the end		
SAHD-6.1	user after initial configuration?	Yes	See Note 15
	Does this include restricting certain system or user		
	accounts, such as service technicians, to least privileged		
SAHD-6.2	access?	N/A	_
	Are all shared resources (e.g., file shares) which are not		
SAHD-7	required for the intended use of the device disabled?	Yes	_

	Are all communication ports and protocols that are not	
SAHD-8	required for the intended use of the device disabled?	Yes
	Are all services (e.g., telnet, file transfer protocol [FTP],	
	internet information server [IIS], etc.), which are not	
	required for the intended use of the device	
SAHD-9	deleted/disabled?	N/A
	Are all applications (COTS applications as well as OS-	
	included applications, e.g., MS Internet Explorer, etc.)	
CALLE 4.0	which are not required for the intended use of the device	
SAHD-10	deleted/disabled?	N/A
	Can the device prohibit boot from uncontrolled or removable media (i.e., a source other than an internal	
SAHD-11	drive or memory component)?	N/A
SAIID II	Can unauthorized software or hardware be installed on	IV/A
SAHD-12	the device without the use of physical tools?	N/A
	Does the product documentation include information	,
SAHD-13	on operational network security scanning by users?	N/A
	Can the device be hardened beyond the default provided	
SAHD-14	state?	N/A
	Are instructions available from vendor for increased	
SAHD-14.1	hardening?	N/A
	Can the system prevent access to BIOS or other	
SHAD-15	bootloaders during boot?	N/A
	Have additional hardening methods not included in	
SAHD-16	2.3.19 been used to harden the device?	N/A

N/A

#### **SECURITY GUIDANCE (SGUD)**

Availability of security guidance for operator and administrator of the device and manufacturer sales and service.

SGUD-1	Does the device include security documentation for the owner/operator?
	Does the device have the capability, and provide
	instructions, for the permanent deletion of data from the
SGUD-2	device or media?
SGUD-3	Are all access accounts documented?
	Can the owner/operator manage password control for all
SGUD-3.1	accounts?
	Does the product include documentation on
SGUD-4	recommended compensating controls for the device?

	Yes	_
•	No	_
	N/A	_
	N/A	_

# HEALTH DATA STORAGE CONFIDENTIALITY (STCF)

The ability of the device to ensure unauthorized access does not compromise the integrity and confidentiality of personally identifiable information stored on the device or removable media.

SICF-1	Can the device encrypt data at rest?
STCF-1.1	Is all data encrypted or otherwise protected?
STCF-1.2	Is the data encryption capability configured by default?
	Are instructions available to the customer to configure
STCF-1.3	encryption?
STCF-2	Can the encryption keys be changed or configured?
STCF-3	Is the data stored in a database located on the device?
STCF-4	Is the data stored in a database external to the device?

TXCF-1

TXCF-2

TXCF-2.1

#### TRANSMISSION CONFIDENTIALITY (TXCF)

The ability of the device to ensure the confidentiality of transmitted personally identifiable information.

Can personally identifiable information be transmitted only via a point-to-point dedicated cable?

Is personally identifiable information encrypted prior to transmission via a network or removable media?

If data is not encrypted by default, can the customer configure encryption options?

N/A	_
Yes	_
N/A	_

Clarius Mobile Health	Clarius Ultrasound App	51-10-00002	22-Oct-2020
TXCF-3 TXCF-4	Is personally identifiable information transmission restricted to a fixed list of network destinations?  Are connections limited to authenticated systems?  Are secure transmission methods	No No	_ _
TXCF-5	supported/implemented (DICOM, HL7, IEEE 11073)?	No	_
	TRANSMISSION INTEGRITY (TXIG) The ability of the device to ensure the integrity of transmitted data.		
	Does the device support any mechanism (e.g., digital signatures) intended to ensure data is not modified		
TXIG-1	during transmission?  Does the device include multiple sub-components	Yes	_
TXIG-2	connected by external cables?	N/A	_
	REMOTE SERVICE (RMOT)		
	Remote service refers to all kinds of device maintenance activities performed by a service person via network or other remote connection.		
RMOT-1	Does the device permit remote service connections for device analysis or repair?	No	_
RMOT-1.1	Does the device allow the owner/operator to initiative remote service sessions for device analysis or repair?	N/A	_
RMOT-1.2	Is there an indicator for an enabled and active remote session?	N/A	_
RMOT-1.3	Can patient data be accessed or viewed from the device during the remote session?	N/A	_
RMOT-2	Does the device permit or use remote service connections for predictive maintenance data?	No	_
RMOT-3	Does the device have any other remotely accessible functionality (e.g. software updates, remote training)?	No	_

## OTHER SECURITY CONSIDERATIONS (OTHR)

NONE

#### Notes:

https://clarius.com/wp-content/uploads/2020/07/MKTG-00046-WP-Rev2.pdf Note 1 Once the Clarius Ultrasound App successfully stores the image remotely (i.e. to the Clarius Cloud or DICOM) the exam is deleted from the device within 10 days, by Note 2  $default. \ This \ time frame \ can \ be \ configurable \ per \ device.$ The Clarius Ultrasound App can import PII from a DICOM Worklist and export PII to a DICOM Storage Server or to the Clarius Cloud. The App communicates with these Note 3 external services via Wi-Fi or Cellular networks. PII can be displayed on the screen of the mobile device Note 4 that is currently running the Clarius Ultrasound App. Images can be exported to Camera Roll (on iOS devices) Note 5 or to Disk (on Android devices). PII can be imported by scanning barcodes that encode Note 6 PII. A proprietary REST API is used to send exams (which Note 7 include PII) to the Clarius Cloud for storage. Clarius Ultrasound App runs on Smart Devices running IOS or Android. Those devices can be configured to force Note 8 login after a predetermined period of time. Note 9 Yes for Smart Device. No for Clarius App. Administrative activities such as Scanner and User admin Note 10 should be done through the Clarius Cloud.

The Clarius Scanner includes an operating system that can be upgraded by the Clarius Ultrasound App. The Clarius Ultrasound App can in turn be upgraded via the mobile device's respective App Store (e.g. Apple App Store or Google Play). The user is responsible for maintaining upgrades to the mobile device itself.

The mobile devices (iOS, Anrdoid) can be configured to auto-update it has installed. This configuration applies to

Note 12 the Clarius Ultrasound App.

The Clarius Ultrasound App doesn't have anti-malware capabilities. It is up to the host mobile device to provide such capabilities. Clarius is not aware of any restrictions for installing security components that may affect the regular use of Clarius Ultrasound App on mobile devices. The Clarius Ultrasound App can communicate with Clarius Scanners by having the scanner act as a wireless access point.

Note 14 access point.

Note 15 Users and passwords are managed through Clarius Cloud.

Physical locks may be managed by the Smart Device

Note 16 owner

Note 11

Note 13

Clarius relies on iOS and Android App sandboxing for

Note 17 storing PHI, as long as the device is not rooted.

On smart devices powered by Apple iOS, Clarius relies on Apple's encryption protection. This feature can't be disabled. On smart devices powered by Android, Clarius recommends that users enable encryption of persistent storage. As rooting may break Android-enforced protection, Clarius recommends that Android users do

Note 18 not use rooted devices.